

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 01/14/2020

OPERATOR NAME:  
ROBERT BAGNAROL  
Permit #: 43570  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:07
BLANK TEST	0.00	07:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11

