

SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 01/02/2020

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:34
INTERNAL STANDARD	VERIFIED	01:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:35
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37

