

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 01/20/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:31
INTERNAL STANDARD	VERIFIED	15:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:32
BLANK TEST	0.00	15:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:34
BLANK TEST	0.00	15:35

