

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 01/06/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:13
INTERNAL STANDARD	VERIFIED	13:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:13
BLANK TEST	0.00	13:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:14
BLANK TEST	0.00	13:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:16
BLANK TEST	0.00	13:16

