

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 10/12/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:33
INTERNAL STANDARD	VERIFIED	08:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:34
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37

