

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 10/19/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:26
INTERNAL STANDARD	VERIFIED	08:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:28
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:29
BLANK TEST	0.00	08:30

