

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 11/15/2020

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:43
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46

