

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 11/08/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:45
INTERNAL STANDARD	VERIFIED	23:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49

