

SUPERVISOR TEST

CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 12/16/2020

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 19:28 |
| INTERNAL STANDARD | VERIFIED | 19:28 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 19:29 |
| BLANK TEST | 0.00 | 19:30 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 19:30 |
| BLANK TEST | 0.00 | 19:31 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 19:31 |
| BLANK TEST | 0.00 | 19:32 |

