

SUPERVISOR TEST

 CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 12/28/2020

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:06
INTERNAL STANDARD	VERIFIED	08:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:09

