

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 12/09/2020

OPERATOR NAME:
JOHN R DEARMAN
Permit #: 36625
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 14:37 |
| INTERNAL STANDARD | VERIFIED | 14:37 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 14:38 |
| BLANK TEST | 0.00 | 14:39 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 14:39 |
| BLANK TEST | 0.00 | 14:40 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 14:40 |
| BLANK TEST | 0.00 | 14:41 |

