

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/14/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:40
INTERNAL STANDARD	VERIFIED	15:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:40
BLANK TEST	0.00	15:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44

