

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/25/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:03
INTERNAL STANDARD	VERIFIED	16:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:03
BLANK TEST	0.00	16:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:04
BLANK TEST	0.00	16:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:05
BLANK TEST	0.00	16:06

