

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 02/07/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:37
INTERNAL STANDARD	VERIFIED	07:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40

