

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 03/02/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:03
INTERNAL STANDARD	VERIFIED	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07

