

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 03/28/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:31
INTERNAL STANDARD	VERIFIED	15:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:32
BLANK TEST	0.00	15:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:33
BLANK TEST	0.00	15:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:34
BLANK TEST	0.00	15:35

