

SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 03/03/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:11
INTERNAL STANDARD	VERIFIED	00:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14

