

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 03/09/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46

