

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

ate: 03/09/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:28
INTERNAL STANDARD	VERIFIED	12:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:32

