

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 04/01/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:36
INTERNAL STANDARD	VERIFIED	16:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:40

