

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 04/01/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:46
INTERNAL STANDARD	VERIFIED	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:47
BLANK TEST	0.00	23:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50

