

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 04/16/2020

OPERATOR NAME:
ROBERT BAGNAROL
Permit #: 43570
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:49
INTERNAL STANDARD	VERIFIED	14:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:51
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:53

