

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 05/03/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:58
INTERNAL STANDARD	VERIFIED	15:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	16:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01

