

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 06/11/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:14
INTERNAL STANDARD	VERIFIED	17:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:14
BLANK TEST	0.00	17:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:16
BLANK TEST	0.00	17:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:17
BLANK TEST	0.00	17:18

