

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 06/13/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:38
INTERNAL STANDARD	VERIFIED	23:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:41

