

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 06/24/2020

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06

