

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 06/27/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45

