

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 07/01/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:45
INTERNAL STANDARD	VERIFIED	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49

