

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 07/29/2020

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:21
INTERNAL STANDARD	VERIFIED	23:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:22
BLANK TEST	0.00	23:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25

