

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/17/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:49
INTERNAL STANDARD	VERIFIED	16:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:51
BLANK TEST	0.00	16:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:52
BLANK TEST	0.00	16:53

