

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/23/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:08
INTERNAL STANDARD	VERIFIED	16:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:08
BLANK TEST	0.00	16:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:11

