

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/05/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:05
INTERNAL STANDARD	VERIFIED	08:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:08
BLANK TEST	0.00	08:09

