

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 09/03/2020

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:47
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:50
BLANK TEST	0.00	23:51

