

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 09/22/2020

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:32
INTERNAL STANDARD	VERIFIED	18:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:32
BLANK TEST	0.00	18:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:34
BLANK TEST	0.00	18:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:35
BLANK TEST	0.00	18:36

