

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/22/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:48
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52

