

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 03/07/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:15
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19

