

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 03/12/2024

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:28
INTERNAL STANDARD	VERIFIED	23:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:30
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32

