

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 03/27/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:56
INTERNAL STANDARD	VERIFIED	08:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:57
BLANK TEST	0.00	08:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	09:00

