

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 05/02/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:20
INTERNAL STANDARD	VERIFIED	00:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:23
BLANK TEST	0.00	00:24

