

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 07/18/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:57
INTERNAL STANDARD	VERIFIED	23:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:57
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:58
BLANK TEST	0.00	23:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:00

