

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

ate: 07/25/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:47
INTERNAL STANDARD	VERIFIED	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:50
BLANK TEST	0.00	23:51

