

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

te: 08/15/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:21
INTERNAL STANDARD	VERIFIED	15:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:22
BLANK TEST	0.00	15:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:23
BLANK TEST	0.00	15:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25

