

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/23/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:15
INTERNAL STANDARD	VERIFIED	23:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:18
BLANK TEST	0.00	23:19

