

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

te: 09/01/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:22
INTERNAL STANDARD	VERIFIED	23:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:25
BLANK TEST	0.00	23:26

