

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 09/07/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:55
INTERNAL STANDARD	VERIFIED	10:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:56
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:58
BLANK TEST	0.00	10:59

