

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

te: 09/14/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:22
INTERNAL STANDARD	VERIFIED	14:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:24
BLANK TEST	0.00	14:25

