

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 09/26/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:15
INTERNAL STANDARD	VERIFIED	16:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:16
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:17
BLANK TEST	0.00	16:18

