

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 10/17/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:19
INTERNAL STANDARD	VERIFIED	10:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:22
BLANK TEST	0.00	10:23

