

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
FAMASTER dmt: 100196

Date: 11/01/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:24
INTERNAL STANDARD	VERIFIED	16:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:24
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:25
BLANK TEST	0.00	16:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:26
BLANK TEST	0.00	16:27

