

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 11/16/2021

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:44
INTERNAL STANDARD	VERIFIED	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48

